

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
 Data File : VN084885.D
 Acq On : 17 Nov 2024 11:02
 Operator : JC\MD
 Sample : VN1117MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1117MBL01

Quant Time: Nov 18 00:15:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	132101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	232197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	195010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	91915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	96793	50.731	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.460%		
35) Dibromofluoromethane	8.165	113	81009	51.542	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.080%		
50) Toluene-d8	10.565	98	262744	45.382	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 90.760%		
62) 4-Bromofluorobenzene	12.847	95	91450	42.264	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 84.520%		

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
Data File : VN084885.D
Acq On : 17 Nov 2024 11:02
Operator : JC\MD
Sample : VN1117MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1117MBL01

Quant Time: Nov 18 00:15:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

