

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
 Data File : VN084893.D
 Acq On : 17 Nov 2024 14:24
 Operator : JC\MD
 Sample : P4845-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-36-20241112

Quant Time: Nov 18 00:20:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	268537	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	236576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	104156	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	111684	49.247	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.500%		
35) Dibromofluoromethane	8.165	113	90593	49.840	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.680%		
50) Toluene-d8	10.565	98	307212	45.882	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.760%		
62) 4-Bromofluorobenzene	12.847	95	113822	45.485	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	90.960%		
Target Compounds						
44) Trichloroethene	9.359	130	1238	0.663	ug/l	Qvalue 47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
Data File : VN084893.D
Acq On : 17 Nov 2024 14:24
Operator : JC\MD
Sample : P4845-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-36-20241112

Quant Time: Nov 18 00:20:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

