

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
 Data File : VN084895.D
 Acq On : 17 Nov 2024 15:12
 Operator : JC\MD
 Sample : P4845-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-35-20241112

Quant Time: Nov 18 00:21:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	298031	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	241875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	111150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114006	49.894	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.780%
35) Dibromofluoromethane	8.159	113	94489	46.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	10.565	98	314553	42.329	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	84.660%#
62) 4-Bromofluorobenzene	12.847	95	121010	43.572	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.140%
Target Compounds						
44) Trichloroethene	9.347	130	73128	35.291	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
Data File : VN084895.D
Acq On : 17 Nov 2024 15:12
Operator : JC\MD
Sample : P4845-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-35-20241112

Quant Time: Nov 18 00:21:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

