

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
 Data File : VN084899.D
 Acq On : 17 Nov 2024 16:48
 Operator : JC\MD
 Sample : P4845-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-5-20241111

Quant Time: Nov 18 00:23:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	149841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	257696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	230033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103410	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	108596	50.178	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.360%			
35) Dibromofluoromethane	8.159	113	89690	51.419	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.840%			
50) Toluene-d8	10.565	98	296575	46.157	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 92.320%			
62) 4-Bromofluorobenzene	12.847	95	113102	47.099	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 94.200%			
Target Compounds						
30) Chloroform	7.965	83	8713	2.593	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
Data File : VN084899.D
Acq On : 17 Nov 2024 16:48
Operator : JC\MD
Sample : P4845-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-5-20241111

Quant Time: Nov 18 00:23:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

