

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
 Data File : VN084900.D
 Acq On : 17 Nov 2024 17:12
 Operator : JC\MD
 Sample : P4845-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-12R-20241111

Quant Time: Nov 18 00:23:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	150559	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	265944	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	105603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	111794	51.409	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.820%			
35) Dibromofluoromethane	8.165	113	92578	51.429	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.860%			
50) Toluene-d8	10.565	98	309304	46.645	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 93.280%			
62) 4-Bromofluorobenzene	12.847	95	117654	47.475	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 94.940%			
Target Compounds						
30) Chloroform	7.959	83	1795	0.532	ug/l #	72
38) Carbon Tetrachloride	8.353	117	12556	4.500	ug/l	89
44) Trichloroethene	9.347	130	22670	12.260	ug/l	97
64) Tetrachloroethene	11.106	164	966	0.612	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
Data File : VN084900.D
Acq On : 17 Nov 2024 17:12
Operator : JC\MD
Sample : P4845-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-12R-20241111

Quant Time: Nov 18 00:23:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

