

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
 Data File : VN084902.D
 Acq On : 17 Nov 2024 18:01
 Operator : JC\MD
 Sample : P4845-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-17-20241111

Quant Time: Nov 18 03:24:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	155704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	266504	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	240107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	106422	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	111277	49.481	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.960%
35) Dibromofluoromethane	8.165	113	92227	51.126	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.260%
50) Toluene-d8	10.565	98	310279	46.693	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.380%
62) 4-Bromofluorobenzene	12.847	95	118060	47.538	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.080%
Target Compounds						
32) 1,1,1-Trichloroethane	8.165	97	12434	3.913	ug/l	92
44) Trichloroethene	9.347	130	8460	4.566	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
Data File : VN084902.D
Acq On : 17 Nov 2024 18:01
Operator : JC\MD
Sample : P4845-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-17-20241111

Quant Time: Nov 18 03:24:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

