

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
 Data File : VN084904.D
 Acq On : 17 Nov 2024 18:49
 Operator : JC\MD
 Sample : P4845-19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-GEP-2-20241111

Quant Time: Nov 18 00:25:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	153449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	261859	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237602	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	110127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	112346	50.690	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.380%			
35) Dibromofluoromethane	8.159	113	90371	50.986	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.980%			
50) Toluene-d8	10.565	98	313048	47.946	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.900%			
62) 4-Bromofluorobenzene	12.847	95	117421	48.120	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.240%			
Target Compounds						
27) cis-1,2-Dichloroethene	7.483	96	854	0.408	ug/l	75
30) Chloroform	7.965	83	1076	0.313	ug/l	96
32) 1,1,1-Trichloroethane	8.159	97	38022	12.140	ug/l	96
44) Trichloroethene	9.347	130	341480	187.560	ug/l	100
64) Tetrachloroethene	11.100	164	7357	4.655	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
Data File : VN084904.D
Acq On : 17 Nov 2024 18:49
Operator : JC\MD
Sample : P4845-19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-GEP-2-20241111

Quant Time: Nov 18 00:25:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

