

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
 Data File : VN084906.D
 Acq On : 17 Nov 2024 19:37
 Operator : JC\MD
 Sample : P4845-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-23-20241112

Quant Time: Nov 18 00:26:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	269354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	237356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	103680	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	112107	50.124	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.240%	
35) Dibromofluoromethane	8.165	113	91439	50.153	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.300%	
50) Toluene-d8	10.565	98	310621	46.250	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	92.500%	
62) 4-Bromofluorobenzene	12.847	95	112626	44.871	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	89.740%	
Target Compounds						
44) Trichloroethene	9.359	130	1056	0.564	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111724\
Data File : VN084906.D
Acq On : 17 Nov 2024 19:37
Operator : JC\MD
Sample : P4845-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-23-20241112

Quant Time: Nov 18 00:26:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

