

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111725\
 Data File : VN088301.D
 Acq On : 17 Nov 2025 14:00
 Operator : JC\MD
 Sample : Q3628-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Nov 18 03:01:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	187868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	422629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	437002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	204472	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	178933	55.080	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.160%
35) Dibromofluoromethane	8.147	113	139109	48.995	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	10.547	98	503393	46.014	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.020%
62) 4-Bromofluorobenzene	12.829	95	185775	48.083	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.160%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	49541	30.355	ug/l	96
18) Methyl Acetate	5.024	43	5133	1.536	ug/l	94
20) Methylene Chloride	5.283	84	17657	6.445	ug/l	88
43) Isopropyl Acetate	8.671	43	23862	2.903	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111725\
Data File : VN088301.D
Acq On : 17 Nov 2025 14:00
Operator : JC\MD
Sample : Q3628-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Nov 18 03:01:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

