

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111725\
 Data File : VN088302.D
 Acq On : 17 Nov 2025 14:21
 Operator : JC\MD
 Sample : Q3647-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-212

Quant Time: Nov 18 03:01:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	188952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	414135	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	410032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	197276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	171725	52.557	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.120%
35) Dibromofluoromethane	8.153	113	105882	38.057	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	76.120%
50) Toluene-d8	10.547	98	469946	43.837	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.680%
62) 4-Bromofluorobenzene	12.829	95	167605	44.270	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.540%
Target Compounds						
16) Acetone	4.430	43	245688	149.674	ug/l	99
20) Methylene Chloride	5.259	84	17897	6.495	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111725\
Data File : VN088302.D
Acq On : 17 Nov 2025 14:21
Operator : JC\MD
Sample : Q3647-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-212

Quant Time: Nov 18 03:01:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

