

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111819\
 Data File : VN059281.D
 Acq On : 18 Nov 2019 14:16
 Operator : JC/SP
 Sample : K5676-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-01-111519

Quant Time: Nov 19 01:10:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	428319	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	659003	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	589186	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	231580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	254245	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
35) Dibromofluoromethane	7.57	113	202745	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
50) Toluene-d8	10.08	98	814756	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
62) 4-Bromofluorobenzene	12.40	95	266383	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
Target Compounds						
16) Acetone	3.80	43	31135	11.372	ug/l	95
20) Methylene Chloride	4.53	84	66875	14.020	ug/l	99
43) Isopropyl Acetate	8.15	43	88490	8.100	ug/l #	92
95) Naphthalene	15.13	128	322755	28.302	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111819\
Data File : VN059281.D
Acq On : 18 Nov 2019 14:16
Operator : JC/SP
Sample : K5676-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-01-111519

Quant Time: Nov 19 01:10:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

