

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111819\
 Data File : VN059289.D
 Acq On : 18 Nov 2019 17:12
 Operator : JC/SP
 Sample : K5900-08 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30698

Quant Time: Nov 19 08:37:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	496626	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	782845	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	692566	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	302788	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	285663	46.79	ug/l	0.00
Spiked Amount			Recovery	=	93.58%	
35) Dibromofluoromethane	7.57	113	229102	49.01	ug/l	0.00
Spiked Amount			Recovery	=	98.02%	
50) Toluene-d8	10.08	98	956912	51.97	ug/l	0.00
Spiked Amount			Recovery	=	103.94%	
62) 4-Bromofluorobenzene	12.40	95	332688	50.11	ug/l	0.00
Spiked Amount			Recovery	=	100.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.75	59	2693819	2315.381	ug/l	99
16) Acetone	3.79	43	188652	59.425	ug/l	94
17) Carbon Disulfide	4.03	76	22280	1.549	ug/l #	92
25) 2-Butanone	6.81	43	35614	8.512	ug/l	97
31) Cyclohexane	7.63	56	188114	19.793	ug/l	94
39) Methylcyclohexane	9.07	83	223473	26.315	ug/l	99
40) Benzene	8.02	78	13813156	626.872	ug/l #	12
43) Isopropyl Acetate	8.14	43	28237	2.176	ug/l #	61
52) Toluene	10.15	92	8066086	610.737	ug/l #	59
67) Ethyl Benzene	11.51	91	1011210	42.216	ug/l	99
68) m/p-Xylenes	11.61	106	1504511	167.677	ug/l	96
69) o-Xylene	11.95	106	482460	55.639	ug/l	99
73) Isopropylbenzene	12.25	105	47241	2.145	ug/l	100
78) n-propylbenzene	12.59	91	43334	1.718	ug/l	99
80) 1,3,5-Trimethylbenzene	12.73	105	50337	2.697	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	98714	5.467	ug/l	97
95) Naphthalene	15.13	128	11149	3.712	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111819\
Data File : VN059289.D
Acq On : 18 Nov 2019 17:12
Operator : JC/SP
Sample : K5900-08 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30698

Quant Time: Nov 19 08:37:28 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

