

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111819\
 Data File : VN059295.D
 Acq On : 18 Nov 2019 19:24
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 19 05:24:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.504	0.445	11.7	71	0.00
3 P	Chloromethane	0.685	0.603	12.0	73	0.00
4 C	Vinyl Chloride	0.664	0.626	5.7#	76	0.00
5 T	Bromomethane	0.382	0.320	16.2	68	0.00
6 T	Chloroethane	0.393	0.385	2.0	80	0.00
7 T	Trichlorofluoromethane	0.794	0.807	-1.6	80	0.00
8 T	Diethyl Ether	0.323	0.364	-12.7	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.479	0.495	-3.3	83	0.00
10 T	Methyl Iodide	0.593	0.592	0.2	77	0.00
11 T	Tert butyl alcohol	0.117	0.148	-26.5#	101	0.00
12 CM	1,1-Dichloroethene	0.466	0.493	-5.8#	84	0.00
13 T	Acrolein	0.060	0.053	11.7	70	0.00
14 T	Allyl chloride	0.958	0.978	-2.1	82	0.00
15 T	Acrylonitrile	0.292	0.346	-18.5	90	0.00
16 T	Acetone	0.320	0.283	11.6	64	0.00
17 T	Carbon Disulfide	1.448	1.265	12.6	72	0.00
18 T	Methyl Acetate	0.759	0.888	-17.0	92	0.00
19 T	Methyl tert-butyl Ether	1.606	1.901	-18.4	92	0.00
20 T	Methylene Chloride	0.557	0.603	-8.3	87	0.00
21 T	trans-1,2-Dichloroethene	0.501	0.531	-6.0	84	0.00
22 T	Diisopropyl ether	1.810	1.996	-10.3	85	0.00
23 T	Vinyl Acetate	1.403	1.683	-20.0	87	0.00
24 P	1,1-Dichloroethane	1.011	1.104	-9.2	86	0.00
25 T	2-Butanone	0.421	0.464	-10.2	82	0.00
26 T	2,2-Dichloropropane	0.884	0.893	-1.0	82	0.00
27 T	cis-1,2-Dichloroethene	0.581	0.651	-12.0	88	0.00
28 T	Bromochloromethane	0.304	0.289	4.9	77	0.00
29 T	Tetrahydrofuran	0.268	0.305	-13.8	87	0.00
30 C	Chloroform	0.970	1.084	-11.8#	89	0.00
31 T	Cyclohexane	0.957	0.899	6.1	76	0.00
32 T	1,1,1-Trichloroethane	0.815	0.905	-11.0	86	0.00
33 S	1,2-Dichloroethane-d4	0.615	0.561	8.8	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.299	0.291	2.7	80	0.00
36 T	1,1-Dichloropropene	0.467	0.486	-4.1	84	0.00
37 T	Ethyl Acetate	0.520	0.584	-12.3	88	0.00
38 T	Carbon Tetrachloride	0.465	0.488	-4.9	85	0.00
39 T	Methylcyclohexane	0.542	0.555	-2.4	80	0.00
40 TM	Benzene	1.407	1.515	-7.7	86	0.00
41 T	Methacrylonitrile	0.226	0.242	-7.1	77	0.00
42 TM	1,2-Dichloroethane	0.483	0.515	-6.6	86	0.00
43 T	Isopropyl Acetate	0.829	0.930	-12.2	89	0.00
44 TM	Trichloroethene	0.346	0.371	-7.2	86	0.00
45 C	1,2-Dichloropropane	0.382	0.419	-9.7#	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111819\
 Data File : VN059295.D
 Acq On : 18 Nov 2019 19:24
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 19 05:24:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.238	0.260	-9.2	90	0.00
47 T	Bromodichloromethane	0.476	0.534	-12.2	89	0.00
48 T	Methyl methacrylate	0.371	0.422	-13.7	89	0.00
49 T	1,4-Dioxane	0.007	0.008	-14.3	97	0.00
50 S	Toluene-d8	1.176	1.084	7.8	75	0.00
51 T	4-Methyl-2-Pentanone	0.469	0.582	-24.1	89	0.00
52 CM	Toluene	0.844	0.935	-10.8#	88	0.00
53 T	t-1,3-Dichloropropene	0.550	0.605	-10.0	87	0.00
54 T	cis-1,3-Dichloropropene	0.591	0.662	-12.0	88	0.00
55 T	1,1,2-Trichloroethane	0.341	0.383	-12.3	91	0.00
56 T	Ethyl methacrylate	0.507	0.630	-24.3	93	0.00
57 T	1,3-Dichloropropane	0.588	0.660	-12.2	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.190	0.226	-18.9	83	0.00
59 T	2-Hexanone	0.360	0.431	-19.7	85	0.00
60 T	Dibromochloromethane	0.361	0.413	-14.4	90	0.00
61 T	1,2-Dibromoethane	0.346	0.391	-13.0	90	0.00
62 S	4-Bromofluorobenzene	0.424	0.418	1.4	81	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.340	0.353	-3.8	84	0.00
65 PM	Chlorobenzene	0.975	1.094	-12.2	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.366	0.415	-13.4	90	0.00
67 C	Ethyl Benzene	1.729	1.987	-14.9#	88	0.00
68 T	m/p-Xylenes	0.648	0.738	-13.9	87	0.00
69 T	o-Xylene	0.626	0.724	-15.7	90	0.00
70 T	Styrene	0.985	1.200	-21.8	90	0.00
71 P	Bromoform	0.284	0.334	-17.6	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.637	4.042	-11.1	89	0.00
74 T	N-amyl acetate	1.541	1.791	-16.2	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.239	1.383	-11.6	92	0.00
76 T	1,2,3-Trichloropropane	1.138	1.217	-6.9	94	0.00
77 T	Bromobenzene	0.914	0.965	-5.6	88	0.00
78 T	n-propylbenzene	4.164	4.694	-12.7	88	0.00
79 T	2-Chlorotoluene	2.583	2.825	-9.4	90	0.00
80 T	1,3,5-Trimethylbenzene	3.082	3.429	-11.3	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.485	-12.5	90	0.00
82 T	4-Chlorotoluene	2.591	2.853	-10.1	88	0.00
83 T	tert-Butylbenzene	2.661	3.005	-12.9	91	0.00
84 T	1,2,4-Trimethylbenzene	2.982	3.422	-14.8	90	0.00
85 T	sec-Butylbenzene	3.461	3.951	-14.2	89	0.00
86 T	p-Isopropyltoluene	3.056	3.561	-16.5	89	0.00
87 T	1,3-Dichlorobenzene	1.565	1.718	-9.8	89	0.00
88 T	1,4-Dichlorobenzene	1.566	1.704	-8.8	89	0.00
89 T	n-Butylbenzene	2.675	3.084	-15.3	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111819\
Data File : VN059295.D
Acq On : 18 Nov 2019 19:24
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 19 05:24:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.614	0.643	-4.7	89	0.00
91 T	1,2-Dichlorobenzene	1.539	1.713	-11.3	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.280	0.284	-1.4	93	0.00
93 T	1,2,4-Trichlorobenzene	0.825	0.966	-17.1	90	0.00
94 T	Hexachlorobutadiene	0.531	0.573	-7.9	91	0.00
95 T	Naphthalene	2.265	2.789	-23.1	95	0.00
96 T	1,2,3-Trichlorobenzene	0.863	0.964	-11.7	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6