

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111820\
 Data File : VN064706.D
 Acq On : 18 Nov 2020 14:14
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Nov 19 00:52:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:09:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	218421	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	449273	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	456839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	197523	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	203344	52.92	ug/l	0.00
Spiked Amount						
			Recovery	=	105.84%	
35) Dibromofluoromethane	7.55	113	136716	53.46	ug/l	0.00
Spiked Amount						
			Recovery	=	106.92%	
50) Toluene-d8	10.06	98	567827	48.86	ug/l	0.00
Spiked Amount						
			Recovery	=	97.72%	
62) 4-Bromofluorobenzene	12.38	95	230908	53.16	ug/l	0.00
Spiked Amount						
			Recovery	=	106.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111820\
Data File : VN064706.D
Acq On : 18 Nov 2020 14:14
Operator : JC/MD
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Nov 19 00:52:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:09:10 2020
Response via : Initial Calibration

