

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111820\
 Data File : VN064719.D
 Acq On : 18 Nov 2020 20:12
 Operator : JC/MD
 Sample : L4783-19 40X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1D

Quant Time: Nov 19 02:47:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:29:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	213482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	438301	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	466081	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	195318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	207469	55.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.48%	
35) Dibromofluoromethane	7.55	113	128089	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
50) Toluene-d8	10.06	98	568032	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
62) 4-Bromofluorobenzene	12.37	95	232112	54.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.54%	
Target Compounds						
11) Tert butyl alcohol	4.75	59	15758	43.250	ug/l	95
19) Methyl tert-butyl Ether	5.00	73	137972	17.965	ug/l	99
52) Toluene	10.13	92	2418	0.312	ug/l	92
68) m/p-Xylenes	11.60	106	754	4.164	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111820\
Data File : VN064719.D
Acq On : 18 Nov 2020 20:12
Operator : JC/MD
Sample : L4783-19 40X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1D

Quant Time: Nov 19 02:47:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:29:54 2020
Response via : Initial Calibration

