


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN111820\
 Data File : VN064698.D
 Acq On : 18 Nov 2020 9:31
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Title : SW846 8260
 Last Update : Wed Nov 18 15:29:54 2020



AutoFind: Scans 3306, 3307, 3308; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	14790	PASS
75	95	30	60	56.6	38360	PASS
95	95	100	100	100.0	67821	PASS
96	95	5	9	6.9	4649	PASS
173	174	0.00	2	1.5	682	PASS
174	95	50	100	67.5	45768	PASS
175	174	5	9	8.8	4023	PASS
176	174	95	101	96.2	44018	PASS
177	176	5	9	8.2	3588	PASS