

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
 Data File : VN069550.D
 Acq On : 18 Nov 2021 14:00
 Operator : JC/MD
 Sample : M4740-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-111721

Quant Time: Nov 19 03:18:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1261609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2220605	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	2214548	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	809459	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	943843	51.597	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.200%			
35) Dibromofluoromethane	8.027	113	650343	48.831	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.660%			
50) Toluene-d8	10.443	98	2826239	53.956	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.920%			
62) 4-Bromofluorobenzene	12.733	95	986127	51.247	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.500%			
Target Compounds						Qvalue
16) Acetone	4.296	43	17254	1.748	ug/l	96
44) Trichloroethene	9.217	130	6257	0.453	ug/l	93
95) Naphthalene	15.501	128	6494m	4.616	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
Data File : VN069550.D
Acq On : 18 Nov 2021 14:00
Operator : JC/MD
Sample : M4740-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-111721

Quant Time: Nov 19 03:18:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

