

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
 Data File : VN069552.D
 Acq On : 18 Nov 2021 14:52
 Operator : JC/MD
 Sample : M4740-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-G-111721

Quant Time: Nov 19 03:18:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1371134	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2378654	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2354924	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	847282	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	996481	50.123	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	100.240%
35) Dibromofluoromethane	8.024	113	694642	48.692	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.380%
50) Toluene-d8	10.443	98	3019712	53.819	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.640%
62) 4-Bromofluorobenzene	12.733	95	1045055	50.701	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.400%
Target Compounds						
16) Acetone	4.299	43	38280	3.569	ug/l	94
20) Methylene Chloride	5.101	84	11526	0.758	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
Data File : VN069552.D
Acq On : 18 Nov 2021 14:52
Operator : JC/MD
Sample : M4740-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-G-111721

Quant Time: Nov 19 03:18:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

