

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
 Data File : VN069553.D
 Acq On : 18 Nov 2021 15:17
 Operator : JC/MD
 Sample : M4740-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-B-111721

Quant Time: Nov 19 03:18:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1349178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2363749	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2356048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	867319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1001617	51.201	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.400%
35) Dibromofluoromethane	8.024	113	680712	48.016	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.040%
50) Toluene-d8	10.443	98	3002495	53.850	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.700%
62) 4-Bromofluorobenzene	12.734	95	1050133	51.268	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.540%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	33968	3.218	ug/l	95
20) Methylene Chloride	5.098	84	12144	0.811	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
Data File : VN069553.D
Acq On : 18 Nov 2021 15:17
Operator : JC/MD
Sample : M4740-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-B-111721

Quant Time: Nov 19 03:18:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

