

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN111821\  
 Data File : VN069557.D  
 Acq On : 18 Nov 2021 17:00  
 Operator : JC/MD  
 Sample : M4739-04  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MW-14D

Quant Time: Nov 19 03:19:32 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N110921W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 09 18:18:22 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units     | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-----------|----------|
| -----                       |        |       |          |          |           |          |
| Internal Standards          |        |       |          |          |           |          |
| 1) Pentafluorobenzene       | 8.088  | 168   | 1306087  | 50.00    | ug/l      | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.971  | 114   | 2304257  | 50.00    | ug/l      | 0.00     |
| 63) Chlorobenzene-d5        | 11.747 | 117   | 2298777  | 50.00    | ug/l      | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.678 | 152   | 846693   | 50.00    | ug/l      | 0.00     |
| System Monitoring Compounds |        |       |          |          |           |          |
| 33) 1,2-Dichloroethane-d4   | 8.440  | 65    | 981892   | 51.85    | ug/l      | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | = 103.70% |          |
| 35) Dibromofluoromethane    | 8.024  | 113   | 674287   | 48.79    | ug/l      | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | = 97.58%  |          |
| 50) Toluene-d8              | 10.443 | 98    | 2932306  | 53.95    | ug/l      | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | = 107.90% |          |
| 62) 4-Bromofluorobenzene    | 12.734 | 95    | 1024122  | 51.29    | ug/l      | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | = 102.58% |          |
| Target Compounds            |        |       |          |          |           |          |
| 16) Acetone                 | 4.296  | 43    | 58747    | 5.75     | ug/l      | 95       |
| 40) Benzene                 | 8.464  | 78    | 20346    | 0.34     | ug/l #    | 88       |
| -----                       |        |       |          |          |           |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN111821\  
Data File : VN069557.D  
Acq On : 18 Nov 2021 17:00  
Operator : JC/MD  
Sample : M4739-04  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MW-14D

Quant Time: Nov 19 03:19:32 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N110921W.M  
Quant Title : SW846 8260  
QLast Update : Tue Nov 09 18:18:22 2021  
Response via : Initial Calibration

