

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
 Data File : VN069561.D
 Acq On : 18 Nov 2021 18:42
 Operator : JC/MD
 Sample : M4746-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-1

Quant Time: Nov 19 03:20:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	1386358	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2413794	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2408461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	920092	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	999327	49.71	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.42%
35) Dibromofluoromethane	8.024	113	695828	48.06	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.12%
50) Toluene-d8	10.443	98	3046230	53.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.00%
62) 4-Bromofluorobenzene	12.733	95	1080912	51.68	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.36%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	28262	2.61	ug/l	98
24) 1,1-Dichloroethane	6.388	63	50599	1.86	ug/l	99
27) cis-1,2-Dichloroethene	7.332	96	117263	7.54	ug/l	97
44) Trichloroethene	9.217	130	386406	25.75	ug/l	93
78) n-propylbenzene	12.924	91	39880	0.48	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
Data File : VN069561.D
Acq On : 18 Nov 2021 18:42
Operator : JC/MD
Sample : M4746-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-1

Quant Time: Nov 19 03:20:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

