

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
 Data File : VN069573.D
 Acq On : 19 Nov 2021 1:07
 Operator : JC/MD
 Sample : PB140846TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB140846TB

Quant Time: Nov 19 03:23:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1297405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2282744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2362531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	897884	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	969396	51.532	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.060%	
35) Dibromofluoromethane	8.024	113	667862	48.781	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.560%	
50) Toluene-d8	10.443	98	2951138	54.807	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 109.620%	
62) 4-Bromofluorobenzene	12.733	95	1063278	53.752	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.500%	
Target Compounds						
						Qvalue
16) Acetone	4.293	43	230250	22.685	ug/l	95
20) Methylene Chloride	5.093	84	40756	2.832	ug/l	85
25) 2-Butanone	7.345	43	67226	4.891	ug/l	93
29) Tetrahydrofuran	7.702	42	27811	3.297	ug/l #	87
43) Isopropyl Acetate	8.566	43	160338	3.845	ug/l #	88
70) Styrene	12.294	104	6172	1.632	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
Data File : VN069573.D
Acq On : 19 Nov 2021 1:07
Operator : JC/MD
Sample : PB140846TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB140846TB

Quant Time: Nov 19 03:23:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

