

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
 Data File : VN069577.D
 Acq On : 19 Nov 2021 2:49
 Operator : JC/MD
 Sample : M4731-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-1600-FLOOD-WALL-JG

Quant Time: Nov 19 05:15:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 19 04:51:32 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1371705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2432946	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2409017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	888661	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	1018630	51.216	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.440%
35) Dibromofluoromethane	8.024	113	553902	37.960	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	75.920%
50) Toluene-d8	10.443	98	3052584	53.191	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.380%
62) 4-Bromofluorobenzene	12.734	95	1073685	50.927	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.860%
Target Compounds						
						Qvalue
16) Acetone	4.299	43	132804	12.376	ug/l	99
18) Methyl Acetate	4.865	43	45716	1.311	ug/l #	76
20) Methylene Chloride	5.101	84	62749	4.124	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111821\
Data File : VN069577.D
Acq On : 19 Nov 2021 2:49
Operator : JC/MD
Sample : M4731-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-1600-FLOOD-WALL-JG

Quant Time: Nov 19 05:15:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 19 04:51:32 2021
Response via : Initial Calibration

