

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
 Data File : VN075314.D
 Acq On : 18 Nov 2022 15:54
 Operator : JC\MD
 Sample : PB149142ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149142ZHE#01

Quant Time: Nov 19 05:07:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	148578	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	262096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	244701	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	94993	47.451	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.900%
35) Dibromofluoromethane	8.171	113	81675	53.083	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.160%
50) Toluene-d8	10.571	98	313567	52.528	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.060%
62) 4-Bromofluorobenzene	12.847	95	103597	49.403	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.800%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	49543	58.447	ug/l	97
20) Methylene Chloride	5.289	84	15362	8.128	ug/l	91
37) Ethyl Acetate	7.571	43	21108	8.080	ug/l	98
43) Isopropyl Acetate	8.694	43	96187	23.737	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
Data File : VN075314.D
Acq On : 18 Nov 2022 15:54
Operator : JC\MD
Sample : PB149142ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149142ZHE#01

Quant Time: Nov 19 05:07:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

