

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
 Data File : VN075317.D
 Acq On : 18 Nov 2022 17:05
 Operator : JC\MD
 Sample : PB149142ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149142ZHE#04

Quant Time: Nov 19 05:09:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	153637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	280035	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	261971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	107466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	100482	48.540	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.080%
35) Dibromofluoromethane	8.171	113	85124	51.780	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.560%
50) Toluene-d8	10.565	98	332785	52.176	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.360%
62) 4-Bromofluorobenzene	12.847	95	108414	48.388	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.780%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	49295	56.239	ug/l	95
20) Methylene Chloride	5.289	84	15299	7.828	ug/l	90
37) Ethyl Acetate	7.565	43	21556	7.723	ug/l	99
43) Isopropyl Acetate	8.694	43	102498	23.674	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
Data File : VN075317.D
Acq On : 18 Nov 2022 17:05
Operator : JC\MD
Sample : PB149142ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149142ZHE#04

Quant Time: Nov 19 05:09:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

