

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
 Data File : VN075318.D
 Acq On : 18 Nov 2022 17:29
 Operator : JC\MD
 Sample : N5638-01RE
 Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-3-111522RE

Quant Time: Nov 18 17:46:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	83443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	142683	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	116536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	47874	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	71358	63.469	ug/l	-0.01
Spiked Amount	50.000	Range	74 - 125	Recovery	=	126.940%#
35) Dibromofluoromethane	8.165	113	61177	73.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	146.080%#
50) Toluene-d8	10.565	98	211632	65.123	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	130.240%#
62) 4-Bromofluorobenzene	12.853	95	65626	57.487	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
Data File : VN075318.D
Acq On : 18 Nov 2022 17:29
Operator : JC\MD
Sample : N5638-01RE
Misc : 5.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-3-111522RE

Quant Time: Nov 18 17:46:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

