

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
 Data File : VN075329.D
 Acq On : 18 Nov 2022 21:52
 Operator : JC\MD
 Sample : PB149142ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149142ZHE#03

Quant Time: Nov 19 05:15:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	179864	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	318959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	287024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	128437	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	111998	46.214	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.420%
35) Dibromofluoromethane	8.171	113	97793	52.227	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.460%
50) Toluene-d8	10.565	98	370529	51.005	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.000%
62) 4-Bromofluorobenzene	12.847	95	122884	48.154	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.300%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	52739	51.395	ug/l	90
20) Methylene Chloride	5.271	84	16372	7.156	ug/l #	79
37) Ethyl Acetate	7.565	43	22810	7.175	ug/l	98
43) Isopropyl Acetate	8.694	43	122725	24.887	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
Data File : VN075329.D
Acq On : 18 Nov 2022 21:52
Operator : JC\MD
Sample : PB149142ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149142ZHE#03

Quant Time: Nov 19 05:15:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

