

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
 Data File : VN075330.D
 Acq On : 18 Nov 2022 22:16
 Operator : JC\MD
 Sample : PB149142ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB149142ZHE#05

Quant Time: Nov 19 05:15:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	172365	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	301832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275307	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	120364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	105332	45.355	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.700%
35) Dibromofluoromethane	8.171	113	92678	52.304	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.600%
50) Toluene-d8	10.570	98	350189	50.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.880%
62) 4-Bromofluorobenzene	12.847	95	116482	48.235	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.460%
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	53112	54.010	ug/l	91
20) Methylene Chloride	5.283	84	16429	7.493	ug/l	94
37) Ethyl Acetate	7.571	43	23494	7.809	ug/l	98
43) Isopropyl Acetate	8.694	43	140410	30.089	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111822\
Data File : VN075330.D
Acq On : 18 Nov 2022 22:16
Operator : JC\MD
Sample : PB149142ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB149142ZHE#05

Quant Time: Nov 19 05:15:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

