

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111823\
 Data File : VN080188.D
 Acq On : 18 Nov 2023 16:07
 Operator : JC\MD
 Sample : 05470-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 20 01:27:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	151979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	285651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	298915	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	132349	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	140161	55.854	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.700%
35) Dibromofluoromethane	8.165	113	108871	51.653	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.300%
50) Toluene-d8	10.571	98	370891	48.815	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.620%
62) 4-Bromofluorobenzene	12.853	95	147749	51.483	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111823\
Data File : VN080188.D
Acq On : 18 Nov 2023 16:07
Operator : JC\MD
Sample : 05470-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 20 01:27:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

