

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111823\
 Data File : VN080192.D
 Acq On : 18 Nov 2023 17:43
 Operator : JC\MD
 Sample : 05490-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Nov 20 01:28:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	143922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	265951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	284491	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	127983	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	135129	56.863	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.720%
35) Dibromofluoromethane	8.171	113	102467	52.216	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.440%
50) Toluene-d8	10.571	98	354697	50.141	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.280%
62) 4-Bromofluorobenzene	12.853	95	139514	52.214	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111823\
Data File : VN080192.D
Acq On : 18 Nov 2023 17:43
Operator : JC\MD
Sample : 05490-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Nov 20 01:28:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

