

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111823\
 Data File : VN080194.D
 Acq On : 18 Nov 2023 18:31
 Operator : JC\MD
 Sample : 05449-29RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-SED-01-GRE

Quant Time: Nov 20 01:29:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	151547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	281612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	299637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134298	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	144696	57.826	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.660%
35) Dibromofluoromethane	8.171	113	77021	37.066	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	74.140%#
50) Toluene-d8	10.570	98	378918	50.586	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.180%
62) 4-Bromofluorobenzene	12.853	95	147120	51.999	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.000%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	52667	62.665	ug/l	91
18) Methyl Acetate	5.041	43	3153	1.168	ug/l #	61
20) Methylene Chloride	5.283	84	6746	3.833	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111823\
Data File : VN080194.D
Acq On : 18 Nov 2023 18:31
Operator : JC\MD
Sample : 05449-29RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-SED-01-GRE

Quant Time: Nov 20 01:29:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

