

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111824\
 Data File : VN084915.D
 Acq On : 18 Nov 2024 13:15
 Operator : JC\MD
 Sample : P4857-01 20X
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ETGI-328

Quant Time: Nov 18 13:33:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	284819	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	254696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	115507	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	122557	52.150	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	104.300%	
35) Dibromofluoromethane	8.165	113	98773	51.234	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.460%	
50) Toluene-d8	10.565	98	328133	46.205	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	92.400%	
62) 4-Bromofluorobenzene	12.847	95	122552	46.174	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	92.340%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111824\
Data File : VN084915.D
Acq On : 18 Nov 2024 13:15
Operator : JC\MD
Sample : P4857-01 20X
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ETGI-328

Quant Time: Nov 18 13:33:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

