

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111824\
 Data File : VN084924.D
 Acq On : 18 Nov 2024 16:52
 Operator : JC\MD
 Sample : P4861-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-11-A-202411

Quant Time: Nov 19 00:30:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	312575	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285945	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132563	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	126751	50.060	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	100.120%	
35) Dibromofluoromethane	8.165	113	99846	47.192	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.380%	
50) Toluene-d8	10.565	98	362247	46.479	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	92.960%	
62) 4-Bromofluorobenzene	12.847	95	155743	53.469	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	106.940%	
Target Compounds						
4) Vinyl Chloride	2.506	62	4479	2.066	ug/l #	77
19) Methyl tert-butyl Ether	5.783	73	14639	2.392	ug/l	94
24) 1,1-Dichloroethane	6.565	63	13846	3.585	ug/l	95
27) cis-1,2-Dichloroethene	7.488	96	181372	75.775	ug/l	88
30) Chloroform	7.959	83	3993	1.016	ug/l	92
40) Benzene	8.606	78	11095	1.177	ug/l	97
42) 1,2-Dichloroethane	8.665	62	8263	2.708	ug/l	100
44) Trichloroethene	9.347	130	47544	21.877	ug/l	100
64) Tetrachloroethene	11.100	164	186140	97.865	ug/l	97
91) 1,2-Dichlorobenzene	14.106	146	11726	2.504	ug/l	90
93) 1,2,4-Trichlorobenzene	15.394	180	5039	1.966	ug/l	87
95) Naphthalene	15.641	128	11339	1.478	ug/l #	93
96) 1,2,3-Trichlorobenzene	15.829	180	3125	1.256	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111824\
Data File : VN084924.D
Acq On : 18 Nov 2024 16:52
Operator : JC\MD
Sample : P4861-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-11-A-202411

Quant Time: Nov 19 00:30:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

