

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111824\
 Data File : VN084929.D
 Acq On : 18 Nov 2024 18:52
 Operator : JC\MD
 Sample : P4870-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-735

Quant Time: Nov 19 00:33:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	207953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	376414	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	343818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	161951	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	148226	49.350	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.700%
35) Dibromofluoromethane	8.159	113	118556	46.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	10.565	98	451669	48.124	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.240%
62) 4-Bromofluorobenzene	12.847	95	180413	51.434	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.860%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	18812	19.047	ug/l	95
43) Isopropyl Acetate	8.688	43	208837	34.969	ug/l #	78
64) Tetrachloroethene	11.100	164	4265	1.865	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111824\
Data File : VN084929.D
Acq On : 18 Nov 2024 18:52
Operator : JC\MD
Sample : P4870-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-735

Quant Time: Nov 19 00:33:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

