

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111824\
 Data File : VN084932.D
 Acq On : 18 Nov 2024 20:04
 Operator : JC\MD
 Sample : P4887-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-739

Quant Time: Nov 19 00:35:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	197893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	356866	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	325064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	148428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	144166	50.439	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.880%		
35) Dibromofluoromethane	8.165	113	113987	47.189	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 94.380%		
50) Toluene-d8	10.565	98	421183	47.334	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 94.660%		
62) 4-Bromofluorobenzene	12.847	95	169649	51.014	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.020%		
Target Compounds						Qvalue
16) Acetone	4.430	43	20118	21.762	ug/l	99
20) Methylene Chloride	5.265	84	2601	1.035	ug/l #	55
43) Isopropyl Acetate	8.688	43	282175	50.257	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111824\
Data File : VN084932.D
Acq On : 18 Nov 2024 20:04
Operator : JC\MD
Sample : P4887-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-739

Quant Time: Nov 19 00:35:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

