

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111825\
 Data File : VN088311.D
 Acq On : 18 Nov 2025 12:16
 Operator : JC\MD
 Sample : Q3659-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB

Quant Time: Nov 19 03:33:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	202456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	464299	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	461775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	221071	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	204210	58.331	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.660%
35) Dibromofluoromethane	8.147	113	154736	49.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.220%
50) Toluene-d8	10.547	98	541322	45.040	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.080%
62) 4-Bromofluorobenzene	12.829	95	201215	47.405	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111825\
Data File : VN088311.D
Acq On : 18 Nov 2025 12:16
Operator : JC\MD
Sample : Q3659-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB

Quant Time: Nov 19 03:33:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

