

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111825\
 Data File : VN088314.D
 Acq On : 18 Nov 2025 13:26
 Operator : JC\MD
 Sample : Q3660-01DL 4X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FRAC-TANK-M21012DL

Quant Time: Nov 19 03:34:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	181013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	414617	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	408427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	191561	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	175524	56.076	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.160%
35) Dibromofluoromethane	8.153	113	137837	49.485	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.980%
50) Toluene-d8	10.547	98	485916	45.274	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.540%
62) 4-Bromofluorobenzene	12.829	95	166322	43.880	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.760%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	14574	9.268	ug/l #	85
44) Trichloroethene	9.330	130	2993	1.034	ug/l	66
64) Tetrachloroethene	11.082	164	102203	37.936	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111825\
Data File : VN088314.D
Acq On : 18 Nov 2025 13:26
Operator : JC\MD
Sample : Q3660-01DL 4X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FRAC-TANK-M21012DL

Quant Time: Nov 19 03:34:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

