

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
 Data File : VN052377.D
 Acq On : 19 Nov 2018 11:01
 Operator : MD\SY
 Sample : VN1119WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1119WBL01

Quant Time: Nov 20 07:20:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	301287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	490912	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	396635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	131937	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	212458	49.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.86%	
35) Dibromofluoromethane	7.59	113	160779	50.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.56%	
50) Toluene-d8	10.09	98	611466	49.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.18%	
62) 4-Bromofluorobenzene	12.40	95	172166	42.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.04%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
Data File : VN052377.D
Acq On : 19 Nov 2018 11:01
Operator : MD\SY
Sample : VN1119WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1119WBL01

Quant Time: Nov 20 07:20:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

