

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
 Data File : VN052379.D
 Acq On : 19 Nov 2018 12:15
 Operator : MD\SY
 Sample : PB114901TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114901TB

Quant Time: Nov 20 07:27:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	295290	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	485455	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	403872	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	131561	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	207611	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
35) Dibromofluoromethane	7.59	113	158053	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
50) Toluene-d8	10.09	98	611489	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
62) 4-Bromofluorobenzene	12.40	95	172198	42.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.56	84	12156	2.82	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
Data File : VN052379.D
Acq On : 19 Nov 2018 12:15
Operator : MD\SY
Sample : PB114901TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114901TB

Quant Time: Nov 20 07:27:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

