

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
 Data File : VN052381.D
 Acq On : 19 Nov 2018 13:34
 Operator : MD\SY
 Sample : J5993-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-LEACHATE-AMENDED-WCC-

Quant Time: Nov 20 07:31:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	297815	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	500223	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	406378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	144368	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	215418	51.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.42%	
35) Dibromofluoromethane	7.59	113	154270	47.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.62%	
50) Toluene-d8	10.09	98	619127	48.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.56%	
62) 4-Bromofluorobenzene	12.40	95	177672	42.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.10%	

Target Compounds				Qvalue		
20) Methylene Chloride	4.56	84	8811	1.83	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
Data File : VN052381.D
Acq On : 19 Nov 2018 13:34
Operator : MD\SY
Sample : J5993-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SPLP-LEACHATE-AMENDED-WCC-

Quant Time: Nov 20 07:31:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

