

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
 Data File : VN052393.D
 Acq On : 19 Nov 2018 18:53
 Operator : MD\SY
 Sample : PB114846ZHE#09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB114846ZHE#09

Quant Time: Nov 20 07:54:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	290583	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	487336	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	422953	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	148617	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	209713	51.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.20%	
35) Dibromofluoromethane	7.59	113	162006	51.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.08%	
50) Toluene-d8	10.09	98	623294	50.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.82%	
62) 4-Bromofluorobenzene	12.40	95	184173	45.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.56%	

Target Compounds						Qvalue
20) Methylene Chloride	4.56	84	7735	1.57	ug/l	97
43) Isopropyl Acetate	8.17	43	17810	2.60	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
Data File : VN052393.D
Acq On : 19 Nov 2018 18:53
Operator : MD\SY
Sample : PB114846ZHE#09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB114846ZHE#09

Quant Time: Nov 20 07:54:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

