

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
 Data File : VN052396.D
 Acq On : 19 Nov 2018 20:13
 Operator : MD\SY
 Sample : J5964-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 LR-RBR-200036

Quant Time: Nov 20 08:00:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	302765	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	506959	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	441694	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	165583	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	228117	53.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.70%	
35) Dibromofluoromethane	7.59	113	170250	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
50) Toluene-d8	10.09	98	642293	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	12.40	95	197607	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
Target Compounds						
16) Acetone	3.82	43	4627	3.93	ug/l	99
20) Methylene Chloride	4.55	84	8764	1.77	ug/l	96
43) Isopropyl Acetate	8.17	43	21207	2.98	ug/l #	94
95) Naphthalene	15.13	128	10266	4.26	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
Data File : VN052396.D
Acq On : 19 Nov 2018 20:13
Operator : MD\SY
Sample : J5964-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
LR-RBR-200036

Quant Time: Nov 20 08:00:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration

