

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
 Data File : VN052381.D
 Acq On : 19 Nov 2018 13:34
 Operator : MD\SY
 Sample : J5993-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-LEACHATE-AMENDED-WCC-

Quant Time: Nov 20 07:31:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 15 03:02:37 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	297815	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	500223	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	406378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	144368	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	215418	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
35) Dibromofluoromethane	7.59	113	154270	47.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.62%	
50) Toluene-d8	10.09	98	619127	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
62) 4-Bromofluorobenzene	12.40	95	177672	42.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.10%	
Target Compounds						
20) Methylene Chloride	4.56	84	8811	1.83	ug/l	Qvalue 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN111918\
Data File : VN052381.D
Acq On : 19 Nov 2018 13:34
Operator : MD\SY
Sample : J5993-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SPLP-LEACHATE-AMENDED-WCC-

Quant Time: Nov 20 07:31:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111418W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 15 03:02:37 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN052381.D

Acq: 19 Nov 2018 13:34

Instrument :

MSVOA_N

ClientSampleId :


SPLP-LEACHATE-AMENDED-WCC-

Tgt Ion:168 Resp: 297815

Ion Ratio Lower Upper

168 100

99 62.6 47.9 71.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

150000



100000

50000

0

Time-->

7.60 7.70 7.80



#20

Methylene Chloride

Concen: 1.83 ug/l

RT: 4.56 min Scan# 868

Delta R.T. 0.01 min

Lab File: VN052381.D

Acq: 19 Nov 2018 13:34

Tgt Ion: 84 Resp: 8811


Ion Ratio Lower Upper

84 100

49 130.7 118.2 177.2

51 51.9 36.1 54.1

86 59.8 51.6 77.4



Abundance Ion 83.90 (83.60 to 84.60): VNO

Ion 48.90 (48.60 to 49.60): VNO

Ion 50.90 (50.60 to 51.60): VNO

Ion 85.90 (85.60 to 86.60): VNO

6000

5000

4000

3000


2000


1000

0

Time-->

4.50 4.55 4.60 4.65









#33
 1,2-Dichloroethane-d4
 Concen: Below ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. -0.00 min
 Lab File: VN052381.D
 Acq: 19 Nov 2018 13:34




Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-LEACHATE-AMENDED-WCC-


Tgt Ion: 65 Resp: 215418
 Ion Ratio Lower Upper
 65 100
 67 51.7 0.0 105.2



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. -0.00 min
 Lab File: VN052381.D
 Acq: 19 Nov 2018 13:34

Tgt Ion: 114 Resp: 500223
 Ion Ratio Lower Upper
 114 100
 63 23.4 0.0 45.4
 88 15.5 0.0 31.2









#35
 Dibromofluoromethane
 Concen: 50.00 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. -0.00 min
 Lab File: VN052381.D
 Acq: 19 Nov 2018 13:34




Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-LEACHATE-AMENDED-WCC-


Tgt Ion: 113 Resp: 154270
 Ion Ratio Lower Upper
 113 100
 111 102.9 81.0 121.6
 192 17.0 13.8 20.6



#50
 Toluene-d8
 Concen: N.D. ug/l
 RT: 10.09 min Scan# 2586
 Delta R.T. -0.00 min
 Lab File: VN052381.D
 Acq: 19 Nov 2018 13:34

Tgt Ion: 98 Resp: 619127
 Ion Ratio Lower Upper
 98 100
 100 61.8 49.5 74.3





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN052381.D

Acq: 19 Nov 2018 13:34

Instrument :

MSVOA_N

ClientSampleId :

SPLP-LEACHATE-AMENDED-WCC-


Tgt Ion: 95 Resp: 177672

Ion Ratio Lower Upper

95 100

174 76.3 0.0 156.0




176 73.3 0.0 148.4



Abundance Ion 94.90 (94.60 to 95.60): VN052381.D (-3212) (-)

Ion 173.80 (173.50 to 174.50): VN052381.D (-3212) (-)

Ion 175.80 (175.50 to 176.50): VN052381.D (-3212) (-)



#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 11.41 min Scan# 2996

Delta R.T. -0.00 min

Lab File: VN052381.D

Acq: 19 Nov 2018 13:34


Tgt Ion: 117 Resp: 406378

Ion Ratio Lower Upper

117 100

82 59.4 46.5 69.7



119 31.4 25.7 38.5




Abundance Ion 116.90 (116.60 to 117.60): VN052381.D (-2903) (-)

Ion 82.00 (81.70 to 82.70): VN052381.D (-2903) (-)

Ion 118.90 (118.60 to 119.60): VN052381.D (-2903) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. -0.00 min

Lab File: VN052381.D


Acq: 19 Nov 2018 13:34

Instrument :

MSVOA_N

ClientSampleId :

SPLP-LEACHATE-AMENDED-WCC-




Tgt Ion:152 Resp: 144368

Ion Ratio Lower Upper

152 100

115 60.4 30.3 91.0

150 154.9 0.0 348.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

