

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064742.D
 Acq On : 19 Nov 2020 18:50
 Operator : JC/MD
 Sample : L4783-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GENIE

Quant Time: Nov 20 03:53:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:29:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	203260	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	414073	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	435328	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183082	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	201329	56.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.62%	
35) Dibromofluoromethane	7.55	113	135575	57.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.06%	
50) Toluene-d8	10.06	98	535095	49.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.92%	
62) 4-Bromofluorobenzene	12.37	95	210592	52.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.20%	

Target Compounds


						Qvalue
5) Bromomethane	2.55	94	524	0.244	ug/l #	41
11) Tert butyl alcohol	4.73	59	357	1.029	ug/l #	70
13) Acrolein	3.75	56	1387	17.078	ug/l #	51
16) Acetone	3.77	43	5427	11.210	ug/l #	88
18) Methyl Acetate	4.28	43	520	1.527	ug/l #	81
19) Methyl tert-butyl Ether	5.00	73	7073	0.967	ug/l #	1
20) Methylene Chloride	4.50	84	644	0.243	ug/l #	34
25) 2-Butanone	6.78	43	813	0.595	ug/l #	51
31) Cyclohexane	7.62	56	6403	1.656	ug/l #	17
36) 1,1-Dichloropropene	7.63	75	19280	4.526	ug/l #	53
38) Carbon Tetrachloride	7.62	117	24643	5.443	ug/l #	15
40) Benzene	8.00	78	4318	0.364	ug/l #	85
51) 4-Methyl-2-Pentanone	9.99	43	230	6.183	ug/l #	39
56) Ethyl methacrylate	10.36	69	31	2.681	ug/l #	1
58) 2-Chloroethyl Vinyl ether	9.51	63	29	13.283	ug/l #	49
59) 2-Hexanone	10.57	43	565	8.618	ug/l	67
67) Ethyl Benzene	11.49	91	3851	0.236	ug/l	97
68) m/p-Xylenes	11.59	106	293	4.104	ug/l #	1
70) Styrene	11.93	104	507	2.580	ug/l #	81
71) Bromoform	12.37	173	1606	3.015	ug/l #	1
73) Isopropylbenzene	12.22	105	10226	0.787	ug/l	93
76) 1,2,3-Trichloropropane	12.37	75	122419	31.033	ug/l #	100
78) n-propylbenzene	12.57	91	4376	0.289	ug/l	85
79) 2-Chlorotoluene	12.57	91	4376	0.452	ug/l #	43
81) trans-1,4-Dichloro-2-buten	12.37	75	122419	79.056	ug/l #	5
82) 4-Chlorotoluene	12.57	91	4376	0.424	ug/l #	43
90) Hexachloroethane	13.91	117	612	0.362	ug/l #	18
93) 1,2,4-Trichlorobenzene	14.88	180	111	1.366	ug/l #	12


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064742.D
Acq On : 19 Nov 2020 18:50
Operator : JC/MD
Sample : L4783-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GENIE

Quant Time: Nov 20 03:53:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:29:54 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.62 min Scan# 1830

Delta R.T. 0.00 min

Lab File: VN064742.D

Acq: 19 Nov 2020 18:50

Instrument :

MSVOA_N

ClientSampleId :


GENIE

Tot Ion:168 Resp: 203260

Ion Ratio Lower Upper

168 100

99 70.1 55.7 83.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.62

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70



#5

Bromomethane

Concen: 0.244 ug/l

RT: 2.55 min Scan# 252

Delta R.T. 0.02 min

Lab File: VN064742.D


Acq: 19 Nov 2020 18:50

Tgt Ion: 94 Resp: 524

Ion Ratio Lower Upper

94 100

96 39.3 77.8 116.8#



Abundance Ion 93.90 (93.60 to 94.60): VNO

Ion 95.90 (95.60 to 96.60): VNO

2.55

400


300


200

100

0

Time--> 2.52 2.54 2.56 2.58









#11

Tert butyl alcohol
 Concen: 1.029 ug/l
 RT: 4.73 min Scan# 931
 Delta R.T. -0.00 min
 Lab File: VN064742.D
 Acq: 19 Nov 2020 18:50

Instrument :
 MSVOA_N
 ClientSampled :
 GENIE




Tot Ion: 59 Resp: 357
 Ion Ratio Lower Upper
 59 100
 57 0.0 9.1 13.7#




#13

Acrolein
 Concen: 17.078 ug/l
 RT: 3.75 min Scan# 626
 Delta R.T. 0.19 min
 Lab File: VN064742.D
 Acq: 19 Nov 2020 18:50

Tgt Ion: 56 Resp: 1387
 Ion Ratio Lower Upper
 56 100
 55 34.8 61.8 92.6#





#16

Acetone

Concen: 11.210 ug/l

RT: 3.77 min Scan# 631

Delta R.T. -0.01 min

Lab File: VN064742.D

Acq: 19 Nov 2020 18:50

Instrument :

MSVOA_N

ClientSampled :


GENIE

Tot Ion: 43 Resp: 5427

Ion Ratio Lower Upper




43 100

58 20.1 21.2 31.8#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#18

Methyl Acetate

Concen: 1.527 ug/l

RT: 4.28 min Scan# 791

Delta R.T. 0.00 min

Lab File: VN064742.D


Acq: 19 Nov 2020 18:50

Tgt Ion: 43 Resp: 520

Ion Ratio Lower Upper



43 100


74 13.1 17.6 26.4#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0









#19

Methyl tert-butyl Ether
 Concen: 0.967 ug/l
 RT: 5.00 min Scan# 1013
 Delta R.T. 0.00 min
 Lab File: VN064742.D
 Acq: 19 Nov 2020 18:50

Instrument :
 MSVOA_N
 ClientSampled :
 GENIE




Tot Ion: 73 Resp: 7073
 Ion Ratio Lower Upper
 73 100
 57 81.4 17.3 25.9#




#20

Methylene Chloride
 Concen: 0.243 ug/l
 RT: 4.50 min Scan# 857
 Delta R.T. -0.01 min
 Lab File: VN064742.D
 Acq: 19 Nov 2020 18:50

Tgt Ion: 84 Resp: 644
 Ion Ratio Lower Upper
 84 100
 49 33.2 102.4 153.6#
 51 30.4 29.4 44.2
 86 13.9 51.2 76.8#









#25
2-Butanone
Concen: 0.595 ug/l
RT: 6.78 min Scan# 1567
Delta R.T. -0.01 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50




Instrument :
MSVOA_N
ClientSampled :
GENIE


Tot Ion: 43 Resp: 813
Ion Ratio Lower Upper
43 100
72 0.0 19.7 29.5#



#31
Cyclohexane
Concen: 1.656 ug/l
RT: 7.62 min Scan# 1830
Delta R.T. 0.01 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50

Tgt Ion: 56 Resp: 6403
Ion Ratio Lower Upper
56 100
69 171.7 27.2 40.8#
84 98.7 65.9 98.9









#33
 1,2-Dichloroethane-d4
 Concen: 56.305 ug/l
 RT: 7.99 min Scan# 1943
 Delta R.T. 0.00 min
 Lab File: VN064742.D
 Acq: 19 Nov 2020 18:50




Instrument :
 MSVOA_N
 ClientSampleId :
 GENIE


Tot Ion: 65 Resp: 201329
 Ion Ratio Lower Upper
 65 100
 67 48.8 0.0 99.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2118
 Delta R.T. 0.00 min
 Lab File: VN064742.D
 Acq: 19 Nov 2020 18:50

Tgt Ion: 114 Resp: 414073
 Ion Ratio Lower Upper
 114 100
 63 22.8 0.0 46.0
 88 16.1 0.0 31.2









#35
 Dibromofluoromethane
 Concen: 57.525 ug/l
 RT: 7.55 min Scan# 1807
 Delta R.T. 0.00 min
 Lab File: VN064742.D
 Acq: 19 Nov 2020 18:50

Instrument :
 MSVOA_N
 ClientSampleId :
 GENIE

Tot Ion:113 Resp: 135575
 Ion Ratio Lower Upper
 113 100
 111 104.4 82.2 123.4
 192 15.4 12.7 19.1




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 4.526 ug/l
 RT: 7.63 min Scan# 1831
 Delta R.T. -0.13 min
 Lab File: VN064742.D
 Acq: 19 Nov 2020 18:50

Tgt Ion: 75 Resp: 19280
 Ion Ratio Lower Upper
 75 100
 110 10.4 16.2 48.6#
 77 0.0 24.4 36.6#



Abundance Ion 74.90 (74.60 to 75.60): V
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): V





#38

Carbon Tetrachloride

Concen: 5.443 ug/l

RT: 7.62 min Scan# 1830

Delta R.T. -0.11 min

Lab File: VN064742.D

Acq: 19 Nov 2020 18:50

Instrument :

MSVOA_N

ClientSampled :

GENIE


Tot Ion:117 Resp: 24643

Ion Ratio Lower Upper

117 100

119 5.2 77.8 116.6#




121 0.0 24.9 37.3#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#40

Benzene

Concen: 0.364 ug/l

RT: 8.00 min Scan# 1948

Delta R.T. 0.00 min

Lab File: VN064742.D


Acq: 19 Nov 2020 18:50

Tgt Ion: 78 Resp: 4318

Ion Ratio Lower Upper



78 100


77 31.5 19.2 28.8#



Abundance Ion 78.00 (77.70 to 78.70): V

Ion 77.00 (76.70 to 77.70): V





#50

Toluene-d8

Concen: 49.956 ug/l

RT: 10.06 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN064742.D

Acq: 19 Nov 2020 18:50

Instrument :

MSVOA_N

ClientSampled :


GENIE

Tot Ion: 98 Resp: 535095

Ion Ratio Lower Upper

98 100

100 63.7 50.4 75.6



Abundance Ion 98.00 (97.70 to 98.70): VN064703.D (-2571) (-)

Ion 100.00 (99.70 to 100.70): VN064703.D (-2571) (-)

10.06



300000

200000

100000

0

Time-->



#51

4-Methyl-2-Pentanone

Concen: 6.183 ug/l

RT: 9.99 min Scan# 2566

Delta R.T. 0.04 min

Lab File: VN064742.D


Acq: 19 Nov 2020 18:50

Tgt Ion: 43 Resp: 230

Ion Ratio Lower Upper

43 100

58 0.0 28.5 42.7#



Abundance Ion 43.00 (42.70 to 43.70): VN064703.D (-2539) (-)

Ion 58.00 (57.70 to 58.70): VN064703.D (-2539) (-)

9.99

600

500

400


300


200

100

0

Time-->





#56

Ethyl methacrylate

Concen: 2.681 ug/l

RT: 10.36 min Scan# 2681

Delta R.T. -0.04 min

Lab File: VN064742.D

Acq: 19 Nov 2020 18:50

Instrument :

MSVOA_N

ClientSampled :

GENIE





Tot Ion: 69 Resp: 31

Ion Ratio Lower Upper

69 100

41 100.0 59.0 88.4#

39 371.0 38.4 57.6#



#58

2-Chloroethyl Vinyl ether

Concen: 13.283 ug/l

RT: 9.51 min Scan# 2415

Delta R.T. -0.16 min

Lab File: VN064742.D




Acq: 19 Nov 2020 18:50


Tgt Ion: 63 Resp: 29

Ion Ratio Lower Upper

63 100

106 0.0 20.9 31.3#









#59
2-Hexanone
Concen: 8.618 ug/l
RT: 10.57 min Scan# 2745
Delta R.T. -0.15 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50




Instrument :
MSVOA_N
ClientSampled :
GENIE


Tot Ion: 43 Resp: 565
Ion Ratio Lower Upper
43 100
58 26.5 24.4 73.4



#62
4-Bromofluorobenzene
Concen: 52.600 ug/l
RT: 12.37 min Scan# 3307
Delta R.T. 0.00 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50

Tgt Ion: 95 Resp: 210592
Ion Ratio Lower Upper
95 100
174 69.3 0.0 136.8
176 66.0 0.0 134.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. 0.00 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50




Instrument :
MSVOA_N
ClientSampleId :
GENIE


Tot Ion:117 Resp: 435328
Ion Ratio Lower Upper
117 100
82 59.3 47.0 70.6
119 31.9 25.4 38.2



#67
Ethyl Benzene
Concen: 0.236 ug/l
RT: 11.49 min Scan# 3031
Delta R.T. 0.00 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50

Tgt Ion: 91 Resp: 3851
Ion Ratio Lower Upper
91 100
106 31.5 23.9 35.9









#68
m/p-Xvlenes
Concen: 4.104 ug/l
RT: 11.59 min Scan# 3064
Delta R.T. 0.00 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50




Instrument :
MSVOA_N
ClientSampled :
GENIE


Tot Ion:106 Resp: 293
Ion Ratio Lower Upper
106 100
91 547.1 171.9 257.9#



#70
Styrene
Concen: 2.580 ug/l
RT: 11.93 min Scan# 3169
Delta R.T. -0.01 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50

Tgt Ion:104 Resp: 507
Ion Ratio Lower Upper
104 100
78 79.3 44.1 66.1#
103 51.1 43.7 65.5









#71
Bromoform
Concen: 3.015 ug/l
RT: 12.37 min Scan# 3307
Delta R.T. 0.27 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50

Instrument :
MSVOA_N
ClientSampled :
GENIE

Tot Ion:173 Resp: 1606
Ion Ratio Lower Upper
173 100
175 639.2 24.3 72.9#
254 0.0 0.1 0.1#




Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.32 min Scan# 3601
Delta R.T. 0.00 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50

Tgt Ion:152 Resp: 183082
Ion Ratio Lower Upper
152 100
115 65.9 32.1 96.3
150 157.4 0.0 344.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 0.787 ug/l

RT: 12.22 min Scan# 3260

Delta R.T. 0.00 min

Lab File: VN064742.D

Acq: 19 Nov 2020 18:50

Instrument :

MSVOA_N

ClientSampled :


GENIE

Tot Ion:105 Resp: 10226

Ion Ratio Lower Upper




105 100

120 29.1 12.8 38.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#76

1,2,3-Trichloropropane

Concen: 31.033 ug/l

RT: 12.37 min Scan# 3307

Delta R.T. -0.15 min

Lab File: VN064742.D


Acq: 19 Nov 2020 18:50

Tgt Ion: 75 Resp: 122419

Ion Ratio Lower Upper



75 100


77 1.0 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 77.00 (76.70 to 77.70): V





#78

n-propylbenzene

Concen: 0.289 ug/l

RT: 12.57 min Scan# 3367

Delta R.T. 0.00 min

Lab File: VN064742.D

Acq: 19 Nov 2020 18:50

Instrument :

MSVOA_N

ClientSampled :


GENIE

Tot Ion: 91 Resp: 4376

Ion Ratio Lower Upper




91 100

120 29.2 10.9 32.9



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 120.00 (119.70 to 120.70): V



#79

2-Chlorotoluene

Concen: 0.452 ug/l

RT: 12.57 min Scan# 3367

Delta R.T. -0.08 min

Lab File: VN064742.D


Acq: 19 Nov 2020 18:50

Tgt Ion: 91 Resp: 4376

Ion Ratio Lower Upper



91 100


126 0.0 16.0 48.0#



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 125.90 (125.60 to 126.60): V





#81

trans-1,4-Dichloro-2-butene

Concen: 79.056 ug/l

RT: 12.37 min Scan# 3307

Delta R.T. 0.10 min

Lab File: VN064742.D

Acq: 19 Nov 2020 18:50

Instrument :

MSVOA_N

ClientSampleId :

GENIE


Tot Ion: 75 Resp: 122419

Ion Ratio Lower Upper

75 100

53 0.3 88.8 133.2#




89 0.0 43.1 64.7#



Abundance Ion 74.90 (74.60 to 75.60): VN064742.D (-3183) (-)

Ion 53.00 (52.70 to 53.70): VN064742.D (-3183) (-)

Ion 88.90 (88.60 to 89.60): VN064742.D (-3183) (-)



#82

4-Chlorotoluene

Concen: 0.424 ug/l

RT: 12.57 min Scan# 3367

Delta R.T. -0.18 min

Lab File: VN064742.D


Acq: 19 Nov 2020 18:50

Tgt Ion: 91 Resp: 4376

Ion Ratio Lower Upper



91 100


126 0.0 15.9 47.7#



Abundance Ion 91.00 (90.70 to 91.70): VN064742.D (-3330) (-)

Ion 125.90 (125.60 to 126.60): VN064742.D (-3330) (-)









#90
Hexachloroethane
Concen: 0.362 ug/l
RT: 13.91 min Scan# 3786
Delta R.T. 0.07 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50

Instrument :
MSVOA_N
ClientSampled :
GENIE

Tot Ion:117 Resp: 612
Ion Ratio Lower Upper
117 100
201 0.0 31.9 95.5#



#93
1,2,4-Trichlorobenzene
Concen: 1.366 ug/l
RT: 14.88 min Scan# 4088
Delta R.T. 0.01 min
Lab File: VN064742.D
Acq: 19 Nov 2020 18:50

Tgt Ion:180 Resp: 111
Ion Ratio Lower Upper
180 100
182 0.0 47.3 141.8#
145 0.0 17.5 52.5#

