

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064739.D
 Acq On : 19 Nov 2020 17:38
 Operator : JC/MD
 Sample : L4783-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-11

Quant Time: Nov 20 05:02:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:29:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	211605	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	427133	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	459075	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	245127	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	211526	56.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.64%	
35) Dibromofluoromethane	7.55	113	143264	58.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.86%	
50) Toluene-d8	10.06	98	567720	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
62) 4-Bromofluorobenzene	12.37	95	222189	53.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.60%	
Target Compounds						
16) Acetone	3.78	43	4491	9.950	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064739.D
Acq On : 19 Nov 2020 17:38
Operator : JC/MD
Sample : L4783-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-11

Quant Time: Nov 20 05:02:44 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:29:54 2020
Response via : Initial Calibration

