

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064747.D
 Acq On : 19 Nov 2020 20:50
 Operator : JC/MD
 Sample : L4769-12 40X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRE-1119

Quant Time: Nov 20 05:30:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:29:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	195383	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	437104	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	433695	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	183280	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	203095	59.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.18%	
35) Dibromofluoromethane	7.55	113	138283	55.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.16%	
50) Toluene-d8	10.06	98	529990	46.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.74%	
62) 4-Bromofluorobenzene	12.37	95	222240	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
Target Compounds						
16) Acetone	3.78	43	10425	17.346	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064747.D
Acq On : 19 Nov 2020 20:50
Operator : JC/MD
Sample : L4769-12 40X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRE-1119

Quant Time: Nov 20 05:30:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:29:54 2020
Response via : Initial Calibration

