

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064748.D
 Acq On : 19 Nov 2020 21:14
 Operator : JC/MD
 Sample : L4826-02DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB-2020-WW-000-04DL

Quant Time: Nov 21 08:16:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:29:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	195898	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	406730	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	431374	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	177582	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	200993	58.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.64%	
35) Dibromofluoromethane	7.55	113	140009	60.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.96%	
50) Toluene-d8	10.06	98	581730	55.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.58%	
62) 4-Bromofluorobenzene	12.38	95	205896	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
Target Compounds						
44) Trichloroethene	8.80	130	2784	0.832	ug/l	83
64) Tetrachloroethene	10.60	164	45310	11.848	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064748.D
Acq On : 19 Nov 2020 21:14
Operator : JC/MD
Sample : L4826-02DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB-2020-WW-000-04DL

Quant Time: Nov 21 08:16:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:29:54 2020
Response via : Initial Calibration

