

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064757.D
 Acq On : 20 Nov 2020 3:55
 Operator : JC/MD
 Sample : L4800-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: Nov 20 07:25:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:29:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	299662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	602602	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	631166	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	266727	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	253561	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
35) Dibromofluoromethane	7.55	113	199945	58.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.60%	
50) Toluene-d8	10.06	98	766645	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	12.38	95	294418	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
Target Compounds						
16) Acetone	3.78	43	6200	9.827	ug/l #	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064757.D
Acq On : 20 Nov 2020 3:55
Operator : JC/MD
Sample : L4800-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: Nov 20 07:25:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:29:54 2020
Response via : Initial Calibration

