

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064764.D
 Acq On : 20 Nov 2020 6:43
 Operator : JC/MD
 Sample : L4798-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20BR-20201118

Quant Time: Nov 24 08:09:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 18 15:29:54 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	222383	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	475206	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	508048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	191791	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	213497	54.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.14%	
35) Dibromofluoromethane	7.55	113	137099	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
50) Toluene-d8	10.06	98	564001	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
62) 4-Bromofluorobenzene	12.38	95	219095	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.83	85	61611	22.173	ug/l	100
4) Vinyl Chloride	2.17	62	2067492	644.870	ug/l	100
9) 1,1,2-Trichlorotrifluoroet	3.71	101	14862	5.989	ug/l #	86
12) 1,1-Dichloroethene	3.69	96	31758	13.283	ug/l	99
21) trans-1,2-Dichloroethene	4.99	96	39977	14.985	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	10042083	3414.311	ug/l	79
44) Trichloroethene	8.80	130	2034413	520.393	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064764.D
Acq On : 20 Nov 2020 6:43
Operator : JC/MD
Sample : L4798-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20BR-20201118

Quant Time: Nov 24 08:09:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 18 15:29:54 2020
Response via : Initial Calibration

