

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111920\
 Data File : VN064766.D
 Acq On : 20 Nov 2020 7:31
 Operator : JC/MD
 Sample : L4778-11 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1814

Quant Time: Nov 24 08:19:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 00:39:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	205018	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	423065	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	447397	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	208151	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	211122	58.54	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	117.08%	
35) Dibromofluoromethane		7.55	113	145944	60.61	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	121.22%	
50) Toluene-d8		10.06	98	564920	51.62	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.24%	
62) 4-Bromofluorobenzene		12.38	95	232359	56.80	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	8829	14.977	ug/l	95
27) cis-1,2-Dichloroethene	6.78	96	8094	2.985	ug/l	43
44) Trichloroethene	8.80	130	5553	1.595	ug/l	95
84) 1,2,4-Trimethylbenzene	13.01	105	15306	1.222	ug/l	97
95) Naphthalene	15.10	128	31397	2.684	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111920\
Data File : VN064766.D
Acq On : 20 Nov 2020 7:31
Operator : JC/MD
Sample : L4778-11 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1814

Quant Time: Nov 24 08:19:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N111820W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 00:39:10 2020
Response via : Initial Calibration

